

Histowax[®]



REF
00402-1
00403

INTENDED PURPOSE

Histowax[®] is a paraffin wax intended for use in in vitro diagnostics for impregnation and embedding of histological and cytological human tissue samples. Impregnation and embedding with Histowax[®] enables pathological analysis and is intended for use by trained personnel in pathology laboratories.

USE

Histowax[®] is a paraffin wax with a melting point of 56-58°C used for impregnation and embedding of tissue samples for in vitro diagnostics.

The melted paraffin infiltrates the tissue and solidifies after embedding to form a matrix that prevents distortion of the tissue structure during sectioning.

The tissue is moved directly from the molten paraffin in the infiltration vessel to an embedding station. It is then placed in an embedding mould containing melted paraffin; it solidifies rapidly when cooled and it is ready for sectioning.

SPECIFICATION

Composition

Paraffin	>99 %
Other constituents	<1 %

Properties

Melting point	56-58°C
---------------	---------

INSTRUCTIONS FOR USE

Storage and shelf life

Store at room temperature, 18-25°C. Histowax[®] should also be protected from moisture and dust.

Histowax[®] has a shelf life of 5 years from the date of manufacture. The shelf life after opening the package is the same as in closed packaging. The expiry date is printed on the package label.

Warnings/precautions for safe handling

Not classified as dangerous.

Ensure good ventilation. Liquid wax: Handle with care since the melted wax gets very hot. Wear protective gloves at high temperatures. Wear protective goggles if there is a risk of splashing.

Waste management

Histowax[®] is not classified as hazardous waste. Used Histowax[®] can be disposed of with ordinary waste. Histowax[®] contaminated with clearing agent (Xylene or HistoLab Clear) must be collected and disposed of according to the regulations for Xylene or HistoLab Clear.

Sample material

Histowax[®] can be used for fixated and dehydrated tissue specimens which are to be subjected to histological examination.

Instructions

Impregnation of tissue samples

The melting temperature of the paraffin in the instrument should be set about 2°C above the specified melting point to make it less viscous. Time in dehydration equipment may vary depending on tissue type and thickness. Large sections and manual impregnation require more time.

Sectioning of paraffin-embedded tissue

Sectioning in a microtome is carried out according to established technique with a section thickness setting of 2-5 µm. Stretching of sections is done in a water bath set at approx: +45°C.

ADDITIONAL INFORMATION

Use equipment and reagents suitable for in vitro diagnostics.

All serious incidents that have occurred in connection with the product must be reported to the manufacturer and the competent authority.

VERSION HISTORY

Latest changes:
Storage

SOURCES

Histotechnology A Self Instructional Text, Freida L. Carson
Theory and Practice of Histological Techniques 3rd Edition, Bancroft/Stevens

Version / datum: v2-2022-10

					
Catalog No.	Batch No.	In Vitro Diagnostic Use	Use By	Consult Instructions for Use	Manufacturer

Histolab Products AB

Södra Långebergsgatan 36
SE-436 32 Askim, Sweden
P: +46 31 709 30 30
mail@histolab.se
www.histolab.se

